



PJLA
Calibration
Accreditation #99092

CALIBRATION REPORT

New Certification

Report Number:	
Date Issued:	Oct-08-20
Tape Serial Number:	
Tape Size:	12" - 24" O.D.

About this Calibration

Temperature	68 Degrees F \pm 2 Degrees
Relative Humidity	45% \pm 15%
Tension	Five Pounds
Manufacturer's Tolerance	\pm .001"
Uncertainty of Measurement	(430 + 0.78L) μ in.

Actual Findings and Procedure Numbers

Tape Reads	Over 20.0000"	Pi Tape Mfg Cal Procedure 1
------------	---------------	-----------------------------

Remarks: New Gage

Calibration Equipment

Equipment	Serial #	Test Date	Due Date	NIST #
20" Ring Gage	042199	Sep-01-20	Sep-01-21	5067260001

This calibration is compliant with the requirements of ISO/IEC 17025:2017 and was performed at Pi Tape Texas, LLC using standards whose results are traceable to the International System of Units (SI) through National Institute of Standards and Technology (NIST). The expanded uncertainties were calculated per the GUM and expressed at a 95% confidence interval and a coverage factor of k=2. Statements of conformance are based on measurement results being within the specified tolerances, which are unmodified by the measurement uncertainties. Decision rules as follows:

- In Tolerance - Results within specifications
- Out of Tolerance - Results exceed specifications

Calibration performed in accordance with the Pi Tape Texas, LLC Q.A. Manual dated 05/19/2020 Rev. B.

This calibration report shall not be reproduced, except in full, without the written approval of the issuing laboratory. Results listed relate only to the item being calibrated.

Report Authorized by: _____

END OF REPORT